

**Begutachtungsausschuss der Österreichischen Gesellschaft für
Hygiene, Mikrobiologie und Präventivmedizin**

A-1090 Wien, Kinderspitalg. 15 Tel: ++ 43 1 40160-33031 Fax: ++ 43 1 40160-933000 DVR: 0504661

Firma
Christof Systems GmbH
Plabutscherstrasse 115
A-8051 Graz

Wien, 11.12.2018



The combined microwave-steam disinfection method

SINTION 1.1

was evaluated for its antimicrobial effectiveness

on 29.12.1995 by Univ.Prof.DDr. Marth (Graz),

on 28.5.1999 by Univ.-Prof.Dr. Werner (Schwerin) and

on 30.6.2006 and on 23.1.2007 by Prim.Dr. Bogatzis (Graz).

On the basis of the respective results the board of experts certifies

the suitability of this method

for treatment of waste from the medical area

by means of electrically generated steam and, for sustaining the temperature,

by means of HF-microwave under the condition of compliance with the appertaining

operating and maintenance instruction.

This expertise is valid until 22. March 22


Univ. Prof. Dr. med. G. Wewalka
Vorsitzender des Begutachtungsausschusses