



Certificate of Compliance

Certificate Number: LR 108460-1

Date Issued: March 12, 1997

Issued to: CMB Maschinenbau und Handels GmbH
Plabutscherstrasse 115
A-8051 Graz
Austria
Attention: Mr. Johann Christof
General Manager

The products listed below are eligible to bear the CSA Mark
NOTE: The "NRTL" indicator also appears adjacent to the CSA Mark.

Issued by: Paul Fabry, P. Eng.
Toronto, ON Canada

Signature: _____

PRODUCTS

Microwave Disinfection System, Model SINTION 1.1, rated 3NPE~/400V ac, 50/60Hz, 10000W.

APPLICABLE STANDARDS

CSA Standards:

CSA Std C22.2 No. 151-M1986 - Laboratory Equipment

UL Standards:

UL Std No 1262 - Laboratory Equipment

The "NRTL" indicator adjacent to the CSA Mark signifies that the product has been evaluated to the applicable ANSI/UL Standards, for use in the U.S.. NRTL, i.e. Nationally Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognized to perform certification to U.S. Standards.



Supplement to Certificate of Compliance

Certificate Number: LR 108460-1

Issued to: CMB Maschinenbau und Handels GmbH
Plabutscherstrasse 115
A-8051 Graz
Austria
Attention: Mr. Johann Christof
General Manager

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Issued By: Paul Fabry, P. Eng.
Toronto, ON Canada

Signature

Product Certification History

Revision	Date	Description
LR 108460-1	March 12, 1997	Original Certification